

MINUTES OF MEETING REGARDING GRIEVANCES OF ANNUAL DEVELOPMENT PROGRAM NO.34 MEDICAL GASSES FOR THE YEAR 2020-21 AT WAZIRABAD INSTITUTE OF CARDIOLOGY, WAZIRABAD

With reference to the letter No.5310-11 dated 15.06.2021, a meeting of grievance committee regarding annual development program No.34 Medical Gasses for the year 2020-21 at Wazirabad Institute of Cardiology, Wazirabad was held on 16.06.2021 at 11:30 AM in the office of Chairman grievance committee WIC, Wazirabad. The meeting started in the name of Allah, the most beneficent the most merciful.

Following members attended the meeting.

S.#.	Name	Designation	
01	Dr. Waseem Ashraf	Assistant Prof. of Cardiology WIC, Wazirabad.	Chairman
02	Dr. Amir Iqbal Qureshi	Assistant Prof. of Cardiac Surgery WIC, Wazirabad.	Member
03	Dr. Irfan Majeed	Assistant Prof. of Cardiology WIC, Wazirabad.	Member
04	Dr. Tasneem Ashraf	Assistant Prof. of Cardiac Surgery WIC, Wazirabad.	Member
05	Dr. Waqas Jamil Chattha	Senior Registrar of Cardiology WIC, Wazirabad.	Member
06	Mst. Sonia Arshad	Pharmacist, WIC, Wazirabad.	Member
07	Mst. Mahjabeen	Head Nurse WIC, Wazirabad.	Member

Meeting was chaired by Dr. Waseem Ashraf, Assistant Prof. of Cardiology, WIC, Wazirabad. The representative of the firm(s) were also present who were heard in person. The committee discusses the grievance raised by firms in detail and following decisions were made in the best interest of public:-


SR No.	Company Name	Grievance Submitted by the Firm	Remarks/Decisions of Grievance Committee
01	M/s Total Technology Against M/s Vertex	Dear Sir, This is in reference to the financial opening meeting held for above referred tender inquiry wherein some anomalies are overlooked and may please be reconsidered as follows:- 1. M/s Vertex Medical Did not quote / offer the prices of "Multiple Zone Valve Box" in their financial offer hence offer of Ms Vertex Medical is incomplete therefore as per bidding documents clause No. 29.3 "Incomplete bid shall stand rejected. All items described in the technical proposal must be priced in financial proposal.	On perusal of the bids it transpired that M/s Vertex Medical clearly quoted zone valve for 2 gases including in alarm area with remote sensor in its financial offer.


1

<p>Medical</p>	<p>2. Further, as per bidding documents Knocked down criteria clause No. 5 "Manufacturer should have documentary evidence to the effect that they are the original manufacturer of the quoted product with indication of manufacturing site and location" whereas as per our understanding M/s Dragger is not the OEM of item "Copper Pipe" therefore the same may please be verified and clarified by calling Quality Certification i.e. CE(MDD) attested from third body. As defined above, the offer of M/s Vertex Medical is incomplete & ambiguous as major anomalies are in the offer (technical / financial), therefore it should be declared as non-responsive.</p>	<p>Furthermore, committee also observed that manufacturer have documentary evidence to effect that they are the original manufacturer of the quoted product with indication of manufacturing site and location on page No. 25 and 28 of technical bid. M/s Dragger has also quality certification (EC Certificate) on page No. 131. In the light of above mentioned observations of committee the grievance of the firm is rejected and upheld the decision of technical evaluation committee/purchase committee.</p>
-----------------------	---	---

The meeting ended with vote of thanks from the chair.

Errors and Omissions accepted.



Miss Mahjabeen
 Head Nurse
 WIC, Wazirabad
 Member


Mst. Sonia Arshad
 Pharmacist
 WIC, Wazirabad
 Member









Dr. Waqas Jamil Chattha
Senior Registrar of Cardiology
WIC, Wazirabad
Member



Dr. Tasneem Ashraf
Assistant Prof. of Cardiac Surgery
WIC, Wazirabad
Member



Dr. Irfan Majeed
Assistant Prof. of Cardiology
WIC, Wazirabad
Member



Dr. Amir Iqbal Qureshi
Assistant Prof. of Cardiac Surgery
WIC, Wazirabad
Member



Dr. Waseem Ashraf
Assistant Prof. of Cardiology
WIC, Wazirabad
Chairman